Proposal for Task Force Consideration at the ISSC 2027 Biennial Meeting	1. a. X Growing Area b. ☐ Harvesting/Handling/Distribution c. ☐ Administrative

Submitter	Titan Fan, Ph.D
Affiliation	Beacon Analytical Systems, Inc
Email	titan@beaconkits.com, holly@beaconkits.com
Proposal Subject	Detection of ASP biotoxins in *Mytilus edulis* (Blue Mussel) shellfish by ELISA for Domoic Acid
Specific NSSP Guide Reference	Section IV. Guidance Documents Chapter II. Growing Areas, Table 2.
Text of Proposal/ Requested Action	SLV Proposal supporting the use of Beacon Domoic Acid Plate Kit as fit for purpose as an Approved NSSP Method for quantification of ASP toxins in Marine Biotoxin Monitoring Programs
Public Health Significance	Shellfish consumption can pose a mammal and bird health risk (1) when toxins produced by cyanobacteria present in water and shellfish growing areas, concentrate in shellfish meat due to their filter feeding system. A Closed Status for any growing areas with shellfish tissue levels of ASP of 2 mg/100 g (20 ppm) or more have been established to protect the consumer from exposure (2). The most common clinical signs of acute toxicity are gastrointestinal distress, confusion and neurological symptoms, disorientation, memory loss, coma and death (3). (1). M.Fernanda, F, Mazzillo, C. Pomeroy, J.Kuo, P. Ramondi,R. Prado, M.Silver. 2010. Aquatic Biol. 9:1-12. (2). NSSP Guide for the Control of Molluscan Shellfish: 2015 Rev. Sec.IV Chp. II., p 231. (3). Kathi A. Lefebvre, Alison Robertson, Toxicon, Vol. 56, Issue 2, 15 Aug. 2010, p. 218-230.
Cost Information	The price per sample is eight to nine dollars dependent upon the number of samples tested during one ELISA run, and/or the volume of kits purchased. There is an ELISA Plate Reader requirement. They can range in price from a low cost unit at approximately $2,600 to a higher cost of $15,000 USD unit depending upon complexity.
Action By 2017 Laboratory Committee	Recommended referral of Proposal 17-108 to an appropriate committee as determined by the Conference Chair.
Action By 2017 Task Force I	Recommended adoption of the Laboratory Committee on Proposal 17-108
Action by 2017 General Assembly	Adopted the recommendation of Task Force I on Proposal 17-108.
Action by FDA February 7, 2018	Concurred with Conference action on Proposal 17-108.
Action by 2019 Laboratory Committee	Recommended referral of Proposal 17-108 to an appropriate committee as determined by the Conference Chair.

Action by 2019 Task Force I	Recommended adoption of the Laboratory Committee on Proposal 17-108
Action by 2019 General Assembly	Adopted the recommendation of Task Force I on Proposal 17-108.
Action by FDA February 21, 2020	Concurred with Conference action on Proposal 17-108.
Action by 2023 Laboratory Committee	Recommended referral of Proposal 17-108 to an appropriate committee as determined by the Conference Chair.
Action by 2023 Task Force I	Recommended adoption of the Laboratory Committee on Proposal 17-108
Action by 2023 General Assembly	Adopted the recommendation of Task Force I on Proposal 17-108.
Action by FDA July 7, 2023	Concurred with Conference action on Proposal 17-108.